

TAN BEAMTUBE # 1 9/27/03
POST BAKE-OUT LEAK CHECK

INFICON UL 1000 DRY LEAK DETECTOR
CALIBRATED LEAK: LDS VAC. PRODUCTS INC.
4.15 X 10⁻¹⁰ ATM. CC/SEC HE @ 22° C
SER. # 1954 MODEL SC-10 D
CALIBRATION DATE 3-19-02