

1/13/04

TAN BEAMTUBE # 3
HELICOFLEX GASKET LEAK CHECK.

UL1000 INFICON DRY LEAK DETECTER
SER. # 9000101745 7

VARIAN V250 DRY TURBO CART

CALIBRATED LEAK: LDS VACUUM PRODUCTS
MODEL SC 10D
SER. # 1954
LEAKRATE - 4.15x10^-10 ATM. cc/SEC. HE.
@ 22°C

Dave Anderson 1/13/04

OMEGA ENGINEERING, INC CHART NO. RDX40-RP