

D

C

B

Officially Released to the LBNL Document Control Center
Released: July 26th, 2004 DCC Auth Key: GYP6CKBNIR

- NOTES, UNLESS OTHERWISE SPECIFIED:
1. THIS IS A ULTRA HIGH VACUUM COMPONENT AND REQUIRES SPECIAL ATTENTION TO AVOID CONTAMINATION. REFER TO LBNL ENGINEERING SPECIFICATION #M20013. HIGH VACUUM USE CLEANING.
 2. TOLERANCES SHOWN IN TITLE BLOCK ARE IN INCHES.

REV	AUTHOR	APPROVER	DATE	CHANGE DESCRIPTION
A	rawlins	plate	07-26-20	INITIAL REVISION

UNLESS OTHERWISE SPECIFIED	
TOLERANCES	.X ± .- .XX ± .03 .XXX ± .010
	FRACTIONS: ± - / - -
	ANGLES: ± 1°
	MACH. SURFS.: 63✓ ^{OR BETTER}
REF: ASME Y14.5M-1994, THREADS ARE CLASS 2	
BREAK EDGES .016 MAX. ON MACHINED WORK	
REMOVE BURRS, WELD SPLATTER & LOOSE SCALE	

MAT 2:		NOTES:	
MAT 1: BAR, ROUND, 1-1/2" SST TYPE 304L		NOTES:	
PROJECT NAME:		ERNEST ORLANDO LAWRENCE	
PROJECT NUMBER:		BERKELEY NATIONAL LABORATORY	
		UNIVERSITY OF CALIFORNIA	
		ALS - INSERTION DEVICE BEAMLINES	
		BL 6.0 FRONT END	
		ABSORBER WATER BLOCK	
DO NOT SCALE PRINT		CATEGORY CODE:	
SCALE: 2 : 1		AL 4 160	
THIRD ANGLE PROJECTION		SHEET 1 OF 1	
		SIZE: C	
		DWG NO: 26B583	
		REV A	

(UNIQUE ID NO.:AL-0008-9588)

