



					UNLESS OTHERWISE SPECIFIED		PROJECT NAME:		ERNEST ORLANDO LAWRENCE BERKELEY NATIONAL LABORATORY UNIVERSITY OF CALIFORNIA		
					TOLERANCES		Advanced Light Source		ALS - INSERTION DEVICE BEAMLINES		
					.X±.06 .XX±.02 .XXX±.005		UNIQUE NUMBER: AL-0013-1549		BL 4.0.2 - MONOCHROMATOR		
					FRACTIONS: ± 1/64		DO NOT SCALE PRINT		MIRROR, MASK		
					ANGLES: ± 0.5°/INCH		CATEGORY CODE:		SHEET 1 OF 1		
					MACH. SURFS.: 125✓ OR BETTER		SCALE: 1 : 1		SIZE: B		
					REF: ASME Y14.5M-1994, THREADS ARE CLASS 2		THIRD ANGLE PROJECTION		DWG NO: 26B746		
					BREAK EDGES .016 MAX. ON MACHINED WORK		MIRROR, MASK		REV A		
					REMOVE BURRS, WELD SPLATTER & LOOSE SCALE						
REV	AUTHOR	APPROVER	DATE	CHANGE DESCRIPTION							
A	wingert	wingert	04-07-05	INITIAL REVISION							



- NOTES, UNLESS OTHERWISE SPECIFIED:
1. DIMENSIONS AND TOLERANCES PER ASME Y14.5M-1994. ALL DIMENSIONS ARE IN INCHES. REFERENCE DIMENSIONS ARE SHOWN IN PARENTHESES. NOT ALL HIDDEN LINES ARE SHOWN.
 2. THIS IS AN ULTRA-HIGH VACUUM (UHV) PART AND REQUIRES SPECIAL ATTENTION TO AVOID UHV CONTAMINATION. READ LBNL SPECIFICATION M20013 BEFORE PERFORMING ANY FABRICATION.
 3. PROTECT FINISHED PART BY BAGGING OR SIMILAR METHOD TO PROTECT AND MAINTAIN CLEANLINESS DURING SHIPMENT AND STORAGE. TAKE CARE NOT TO DING, SCRATCH, DENT, OR OTHERWISE MARK ANY ANY MACHINED SURFACE.