



Sara- see
comments on next
page

Ceramtec
Conector

Cables PMTs





How can these fittings be tightened (around the tube) after this manifold is installed? It appears they can only be tightened beforehand. Are you using a flexible tube of some sort? Also, they look too close together to get a wrench on them.

If you had the fittings on this side you might be able to slide them over the tubes, then tighten them up. You would have to use plastic ferrules so the tubes could slide out for servicing. This is what I did on my pressure chamber. Metal ferrules will get locked onto the tube.