


model is the one I received on 23/5/12, it could be out of date, so maybe the design has changed significantly? I moved the open half of the castle to close against the other half. Please send me a STEP file of the new design, when appropriate




10 cm here, do you mean to have 20 cm?


15 cm here, I think this should be 11-12 cm

this pipe not used

tracking cables

same thin as before, do
you intend for these to
overlap?





this aperture appears to be off center(in the longitudinal direction, check drawing N100_D-side-view.pdf of 13/6/12


this pipe not used


this aperture not needed

this pipe not used

are there
wheel guides
here?

wheels do not match
up with I-beam web





We will need threaded holes or other means of fastening the lead inserts to each castle half. Some sort of retaining plate will be needed

cables or pipes

this insert mounts to the left half of the castle

this pipe not used

this insert mounts to the right half of the castle