



LAWRENCE LABORATORY BERKELEY

Mr. Derek Shuman
Mail Stop 46R0125
1 Cyclotron Road
Berkeley, CA. 94720

06/23/2010
Laboratory No: B427
Client No.: New
PO No.: 450824
Page 1 of 1

Mr. Shuman:

SUBJECT: Hydrostatic test on a stainless steel octagon chamber.

TEST: At your request Testing Engineers Inc. has conducted Hydrostatic Pressure Test on one (1) MFC6DO-S0200 800 Octagon Chamber. Utilizing a Richard Dudgeon Test Pump (Model 7-TSS.1 6000 PSI capacity s/n 5453), and a calibrated pressure gage (s/n 7624 0 to 600 PSI).

PROCEDURE: The chamber was placed on a table attached to the hydrostatic pump. The chamber was then filled with water and pressurized until a pressure of 525 PSI was obtained. The test pressure was held for a minimum of five (5) minutes.

RESULTS: No leaks or loss of pressure were observed during the test. The chamber was then visually examined for sign of distress, none were apparent.

NOTE: Testing was witnessed by Mr. Derek Shuman and Mr. Tom Miller of Advanced Light Sources, Lawrence Laboratory Berkeley.

If you should have any questions concerning this test or if we may assist you in any other way please call at 1-510-835-3142 ext. 106.

TESTING ENGINEERS, INC.

By

Jack A. Snow
Materials Testing Technician

Attachments:

The results presented in this report relate only to the item(s) tested.
This report can be reproduced only in its entirety unless written permission from TEI is obtained



2900 Main St Alameda CA 94501 Phone (510)522-8326 Fax (510)522-3136

Certificate of Calibration

TESTING ENGINEER, INC.
2811 TEAGARDEN ST
SAN LEANDRO
CALIFORNIA 94577

Customer ID # 2103

Rated Accuracy 1%

File # 321

Pass/Fail as Found PASS

Instrument Type PRESSURE GAUGE

Pass/Fail as Left PASS

1st (Mfg) S/N 7624

Range 0-600

Units PSIG

Resolution 20

2nd S/N N/A

Mfg. CPI

Model 2.5-600

Cal By R.K. STRAHL

Cal Date 6/21/2010

Cal Due 6/21/2011

Curent Cal Cycle (Months) 12

Previous Cal Cycle 12

Standards Used AMETEK PK-654WC S/N
73272 DUE 1/28/2012 NIST#
35817.001

Notes

TECHNICAL SERVICES GROUP CERTIFIES THAT THIS INSTRUMENT
HAS BEEN CALIBRATED TRACEABLE TO THE NATIONAL INSTITUTE
OF STANDARDS AND TECHNOLOGY AND CONFORMS TO ISO 10012
AND ANSI / NCSL Z-540. UNLESS OTHERWISE SPECIFIED
MEASURMENT UNCERTAINTIES ARE LESS THAN 1/4 OF TOLERANCE
OR 1 MINOR DIVISION.